Generic Contrast Agents Our portfolio is growing to serve you better. Now you have a choice.





This information is current as of May 18, 2025.

Anatomically Conformable, Three-Dimensional, Detachable Platinum Microcoil System for the Treatment of Intracranial Aneurysms

David Niemann, Richard Aviv, Caroline Cowsill, Mary Sneade and Andrew J. Molyneux

AJNR Am J Neuroradiol 2004, 25 (5) 813-818 http://www.ajnr.org/content/25/5/813

Anatomically Conformable, Three-Dimensional, Detachable Platinum Microcoil System for the Treatment of Intracranial Aneurysms

David Niemann, Richard Aviv, Caroline Cowsill, Mary Sneade, and Andrew J. Molyneux

BACKGROUND AND PURPOSE: The configurations of detachable platinum coils have evolved over the last decade. Our objective was to assess the efficacy of the anatomically conformable three-dimensional (ACT) detachable platinum microcoil in the treatment of intracranial aneurysms.

METHODS: A series of 141 patients presented between May 2000 and August 2002 to a single neurointerventionalist for endovascular treatment of intracranial aneurysms. Eight patients were treated at another institute or are now overseas. The remaining 133 patients had 141 aneurysms, which were treated; 110 were ruptured and 31 were unruptured. The clinical conditions and aneurysm locations were recorded. The procedural technique was the same as that used with the Guglielmi detachable coil (GDC) device, though the coil detachment mechanism was unique. The ACT microcoil was used in almost all cases as the initial coil.

RESULTS: Complete (100%) occlusion was achieved in 104 (74%) aneurysms. Thirty four (24%) had subtotal occlusion (\geq 95%), and three (2%) had incomplete occlusion (<95%). The ACT coil conformed to both spherical and irregularly shaped aneurysms. No procedure-related deaths occurred, and there was no aneurysm rupture related to coil deployment. No postprocedural rebleeding was observed.

CONCLUSION: The ACT microcoil device is feasible as an alternative platinum coil device, offering some advantages over other currently available devices.

The use of detachable platinum coils to treat intracranial aneurysms has developed rapidly since 1992, when they were first introduced into clinical practice (1, 2). The original Guglielmi detachable coil (GDC; Target Therapeutics/Boston Scientific, Natick, MA) has undergone multiple evolutions in coil range, size, and configuration to include soft or ultrasoft, twodimensional (2D) or three-dimensional (3D), and stretch-resistant coils to improve the range of aneurysms that can be treated and to promote better

From the Neurovascular Research Unit, Department of Neuroradiology (R.V., C.C., M.S., A.J.M.), The Radcliffe Infirmary, United Kingdom (A.J.M.); Department of Neurological Surgery and Radiology (D.N.), University of Wisconsin Medical School, Madison, WI; and the Micrus Corporation, Mountain View, CA. aneurysm packing (3). We describe a new anatomically conformable 3D (ACT) detachable platinum microcoil (Micrus Corporation, Sunnyvale, CA) with two unique features: 1) a truly spherical, highly conformable configuration of the first coil that assists in the production of a reliable outer basket and 2) a detachment system that detaches the coil from the delivery wire within 5 seconds. We evaluated the use of this device in a consecutive series of 141 patients presenting for endovascular coil treatment of an intracranial aneurysm.

Methods

Received May 13, 2003; accepted after revision October 18.

Presented at the World Interventional Neuroradiology meeting Val d'Isere, 2002, and the American Association of Neurological Surgeons/Congress of Neurological Surgeons Joint Meeting American Society of Interventional and Therapeutic Neuroradiology 2002.

Address reprint requests to Andrew J. Molyneux, MB, BChir, FRCR, Neurovascular Research Unit, Department of Neuroradiology, The Radcliffe Infirmary, Woodstock Road, Oxford OX2 6HE, United Kingdom.

[©] American Society of Neuroradiology

A series of 141 patients with 150 aneurysms presented to a single neurointerventionalist for endovascular aneurysm treatment. Of these patients, 133 were treated at one institution, and eight were treated at another institution or were abroad at the time of follow-up and not included. At initial neurologic evaluation, a World Federation of Neurological Surgeons (WFNS) grade was assigned (4). Both the neurosurgeon (D.N.) and the neurointerventionalist (A.J.M.) completed these examinations. The assigned grades and presentations are shown in Table 1. Patients underwent both CT scanning and digital subtraction angiography to confirm the presence of subarachnoid hemorrhage (SAH) and a cerebral aneurysm. All patients except one were treated with the Micrus 10 system. Clinical outcome measures at follow-up were collected by using the

 TABLE 1: Clinical presentations and grades

Presentation	No. of Patients $(n = 133)$
- Tesentation	(155)
SAH	110
WFNS grade 1*	61
WFNS grade 2*	13
WFNS grade 3*	7
WFNS grade 4*	7
WFNS grade 5*	22
Incidental	15
Mass	3†
Headache	2
Ischemic stroke	1
Seizure	2

* At the time of treatment.

[†] Cranial nerves II and III.

modified Rankin scale (5) (Table 2). These measures were obtained by means of clinical assessment or questionnaire (6). The angiographic outcome was obtained by assessing the appearance of the aneurysm on digital subtraction angiograms or MR angiograms, as reviewed by the treating neurointerventionalist. Occlusion was categorized on the basis of three grades of occlusion, as described in Table 3. We excluded patients with aneurysms greater than 15 mm in luminal diameter for which a 0.018-inch coil device was considered more appropriate; however, one patient was treated with the Micrus 18 system when it became available. Patients randomly selected in the International Subarachnoid Aneurysm Trial (ISAT) (7) were also excluded in the early part of the study, but subsequent approval by the ISAT steering committee allowed us to use the Micrus device in ISAT patients after February 2001.

Description of the Device

The platinum coil device was available as a spherical coil (MicruSphere) for initial placement and as helical (HeliPaq), J coils (InterPaq), and ultrasoft finishing coils (UltiPaq) for packing. The coils were sized 2–10 mm in the Micrus 10 system. The spherical coil, which conforms to both spherical and irregularly shaped aneurysms, was routinely used as the first coil in almost all cases (Figs 1–3).

The coil is detached by fracturing a tensioned polyester fiber after a 5-second current cycle is applied to an insulated, resistive heating coil in the device-positioning unit. The detachment junction zone has some flexibility to allow angulation relative to the tip of the microcatheter.

Procedural Technique

The technique was the same as that used with the GDC device. The relevant cerebral vessel was catheterized with a 6F

TABLE 2: Questionnaire used to assess modified Rankin scale scores

guide catheter. The aneurysm was catheterized with a 0.14-inch or 0.10-inch microcatheter, often a Rebar 14 (Microtherapeutics, Irvine, CA), a Tracker Excel 14 (Target Therapeutics/Boston Scientific, Fremont, CA), or a Prowler 14 (Cordis, Miami Lakes, FL), with appropriate stream shaping of the catheter tip. Catheter position is verified by means of angiography, and the first coil delivered under a subtraction roadmap. The balloon remodeling technique was used in 19 patients. In seven patients, a combination of Micrus ACT coils and GDC coils were used. All patients underwent anticoagulation during the procedure with systemic heparinization to maintain an activated clotting time of about twice the baseline. Heparinization continued for 24 hours after treatment. This regimen was used unless neurosurgical intervention was recently performed or planned to follow embolization.

Results

Patient age range was 25–88 years, with a mean of 54 years. Eighty-one patients were women, and 52 were men. A total of 110 patients presented with SAH; 29 of these patients had a poor clinical grade (WFNS grade of 4 or 5) at the time of the treatment. The remainder had a good clinical grade (WFNS grade of 1–3) after SAH. The locations and sizes of the aneurysms are shown in Table 4. A total of 680 coils were deployed, with a mean of 4.9 coils per aneurysm and a median of 5 coils per aneurysm. In total, 141 aneurysms were treated. The sizes ranged from 2 to 15 mm in lumin diameter.

Clinical Outcomes

Table 5 shows the patients' pretreatment WFNS grades after SAH and their modified Rankin outcome at last follow-up. Outcome measures were obtained by means of clinical assessment or questionnaire (6). Follow-up was 1-29 months, with an average of 15 months. No instances of delayed rebleeding were observed during follow-up. Thirteen patients died; all had a recent SAH. Ten patients with a poor clinical grade at the time of treatment died as a result of the severity of the initial bleed. Regarding the other three patients, one had a clinical grade of 2 with severe hemispheric vasospasm, one died of liver cancer 6 months later, and one died from head trauma. No procedure-related deaths occurred, and there was no aneurysm rupture related to coil deployment. The mortality rate was 8%.

Score	Functional Outcome	Questionnaire Response
0	No symptoms	"I have no symptoms and I cope well with life."
1	Minor symptoms	"I have a few symptoms but these do not interfere with my everyday life."
2	Some restriction in lifestyle	"I have symptoms, which have changed my life but I am still able to look after myself."
3	Significant restriction in lifestyle	"I have symptoms, which have significantly changed my life and prevent me from coping fully, and I need some help looking after myself."
4	Partly dependent	"I have quite severe symptoms, which means I need to have help from other people but I am not so bad as to need attention day and night."
5	Fully dependent	"I have major symptoms, which severely handicap me and I need constant attention day and night."
6	Dead	Not applicable.

 TABLE 3: Aneurysm angiographic outcomes

Occlusion Grade	No. of Aneurysms
At treatment $(n = 141)$	
Ι	104 (74%)
II	34 (24%)
III	3 (2%)
At follow-up $(n = 94)$	
Ι	74 (79%)
II	17 (18%)
III	3 (3%)

Note—Average radiographic follow-up was 10 months. Occlusion grades were as follows: I, complete angiographic occlusion, no neck remnant, approximately 100%; II, subtotal occlusion, coils at the neck but some body filling or small neck remnant, approximately 95–99%; and III, incomplete occlusion, residual aneurysm body filling or sizable neck remnant or coil compaction, <95%.

Angiographic Outcomes

Complete angiographic occlusion (100%) was achieved in 104 (74%) aneurysms. Thirty-four (24%) had subtotal occlusion ($\geq 95\%$), and three (2%) had incomplete occlusion (<95%). These occlusion categories were estimates made by the treating neurointerventionalist. Ninety-four aneurysms were evaluated after radiographic follow-up: 79 with angiography and 15 with MR angiography. Stable occlusion was shown in 88 (95%) aneurysms evaluated with follow-up imaging thus far. There have been five successful repeat treatments: three were endovascular, one involved coils, and two involved a liquid polymer (Onyx; MTI, Irvine, CA). Two patients required surgery. One patient underwent cranitomy and clipping, and one patient required revascularization and trapping of the aneurysm. Angiographic follow-up was 2-22 months, with an average of 10 months. Among the patients still due for radiographic follow-up, 35 aneurysms were pending evaluation. The results are summarized in Table 3.

Procedural Complications

One rupture due to microcatheter perforation occurred in a patient with a small, unruptured pericallosal aneurysm. The aneurysm was immediately treated with coils, and the patient had a good clinical outcome. No coil-related perforation was observed in this series, and there were no intraprocedural adverse clinical events attributable to the coil device. One parent-artery occlusion resulted from overpacking at the anterior communicating artery, which compromised the A2 segments of the anterior cerebral artery bilaterally. However, there was adequate flow from the posterior circulation through the pericallosal collaterals, and the patient's clinical result was given a modified Rankin score of 2 at follow-up.

One coil migrated into the distal right middle cerebral artery when a second coil displaced the first despite balloon remodeling; this occurred during the treatment of a 3-mm internal carotid artery aneurysm. An attempt to remove this coil with a snare was unsuccessful. The patient received anticoagulants for 4 days after the procedure, and antiplatelet therapy was continued for 2 months. No clinical sequelae occurred, and the artery remained open, as observed at exploratory surgery for a posterior communicating aneurysm 1 year later. There was one incidence of coil stretching and fracture; this occurred after excessive manipulation resulted in a coil length being left in the middle cerebral trunk during treatment of a right middle cerebral artery aneurysm. The patient recovered well with a modified Rankin score of 1. Complete occlusion was noted at 6-month follow-up angiography. Two patients with middle cerebral artery aneurysms had thromboembolic complications after coiling. They both developed hemiparesis and aphasia and were given abciximab (Reopro; Eli Lilly, Indianapolis, IN) in the postanesthesia recovery unit. Their deficits resolved within several hours. At follow-up, both patients' clinical outcome was a modified Rankin score of 1. The overall procedure-related morbidity rate was 3%.

Technical Complications

In the early period (in 2000), two cases of partial detachment failure occurred. This type of failure resulted when the coil remained adherent to the device positioning unit after the detachment cycle was performed but later became detached in the microcatheter during retrieval and was redeployed by advancing the positioning unit. In one case, a loop of coil was present in the parent artery without adverse clinical sequelae. In the other case, the patient had a poor clinical grade (WFNS grade 5) after SAH and died from disease-related complications. There was one complete detachment failure with removal of the coil. Occasionally, more than one detachment cycle was

Fig 1. Patient 1.

A, Small, wide-necked, ruptured aneurysm of the anterior cerebral artery. *B*, After treatment with a single, spherical ACT coil.



A, Ruptured aneurysm of the middle cerebral artery bifurcation with a spherical dome and proximal tubular segment.

B, After treatment with a 6-mm, spherical ACT coil and four small helical coils.



Fig 3. Patient 3.

A, Wide-necked, ruptured aneurysm of the posterior communicating artery.

B, After treatment with a 5-mm, spherical ACT coil as a basket followed by several small helical coils.

required to detach the coil. With modifications to the detachment system, this problem is now uncommon.

Discussion

Occlusion of intracranial aneurysms with platinum coils has become an integral part of the management of both ruptured and unruptured aneurysms (7–11). To date, more than 100,000 patients worldwide have undergone endovascular platinum coil treatment, nearly all with the Target Therapeutics/Boston Scientific GDC device (12). The proportion of patients treated with coils varies widely in different countries. ISAT results have shown that this technique has a

TABLE 4: Aneurysm	location	and	size
-------------------	----------	-----	------

	No. of Aneurysms
	(n = 141)
Arterial location	
Internal carotid	47
Anterior cerebral	30
Anterior communicating	15
Middle cerebral	33
Posterior cerebral	1
Superior cerebellar	1
Basilar termination	9
Anterior inferior cerebral	2
Vertebral	3
Size	
<10 mm with neck <4 mm	75
$<10 \text{ mm}$ with neck $\geq 4 \text{ mm}$	40
$\geq 10 \text{ mm}$ with neck $< 4 \text{ mm}$	8
≥ 10 mm with neck ≥ 4 mm	18

major role in selected patients with ruptured aneurysms, improving 1-year clinical outcome in patients with a good clinical grade with small anterior circulation aneurysms suitable for both clipping and coiling (7).

Modified Rankin Score	No. of Patients
SAH, WFNS grades 1–3 pretreatment ($n = 80$)	
0	25
1	35
2	10
3	6
4	1
5	0
6	3
SAH, WFNS grades 4–5 pretreatment $(n = 30)$	
0	3
1	6
2	5
3	4
4	0
5	2
6	10
Unruptured aneurysm $(n = 23)$	
0	10
1	6
2	5
3*	1
4	0
5	1
6	0

Note-Average clinical follow-up was 15 months.

* Patient presented with ischemic stroke.

The technical challenges of small and wide-necked aneurysms have reduced the applicability of the GDC device, as reported in some centers (9, 13). The availability of a wider range of coils with extremely compliant and soft configurations have important applications, particularly in the treatment of small aneurysms. These aneurysms represented more than 90% of patients presenting for treatment after rupture, as shown in the ISAT study (7). The introduction of balloon remodeling has further increased the proportion of intracranial aneurysms that are treatable with endovascular means. This technique is particularly applicable to side-walled aneurysms and, with newer conformable balloons, to certain bifurcation aneurysms.

The newly reported treatments with liquid materials (14) are likely to be applied to larger and widenecked aneurysms first, in which the long-term durability of coil treatment is questionable. The low rate of procedure-related complications with platinum coils in this series and other recent series may be difficult to surpass in the short term (9). Current ongoing studies will help to determine the role of liquid agents.

Patients had aneurysms covering a wide range of shapes and sizes, from small, 2-mm aneurysms to irregular aneurysms as large as 15 mm. The ACT microcoil conformed to both near-spherical and irregularly shaped aneurysms (Figs 1-3). In certain instances, more than one spherical coil was used for separate lobules, and some have advocated the use of a spherical coil within a spherical coil in a Russian doll-type situation. The ACT microcoil seemed particularly useful for small aneurysms with relatively wide necks to provide a stable first basket, either with or without balloon remodeling. Complete or subtotal angiographic occlusion was achieved in all but three patients, and no device-related adverse events occurred. Three of 680 coils (from the first production) had some type of detachment failure. Any detachment problems appear to have been largely resolved. The nature of the junction zone of the coil with the positioning unit is based on a resistive heating coil that is quite flexible and that enables its use around relatively tight bends at the aneurysm neck (eg, small lateral-wall aneurysms) without causing undue catheter displacement. The rapid detachment mechanism is helpful, especially in large aneurysms. In seven patients, both GDC and ACT coils were used because of the lack of a correct helical ACT coil size. However, the coil platinum alloys are virtually identical, and the coils can be used together without difficulty through the same range of microcatheters.

The immediate clinical outcomes were satisfactory and comparable to those of the existing GDC device. Rupture during coil deployment did not occur in any patients in this series. The numbers were too small to determine if our rupture rate differed from the 3-5%associated with GDCs (8–11). There is no reason to suppose that the overall procedural complication rate differs from that of the GDC device.

The current data suggest that endovascular coil treatment is successful in preventing early and mid-

term rebleeding (8, 15). The occlusions achieved with the ACT microcoil device are encouraging, but whether the long-term results differ from those of other available coils is unclear. The ability to achieve excellent acute occlusion by using the spherical first coil may or may not improve the long-term durability of the angiographic occlusion; however, no instances of delayed rebleeding occurred in this patient population during follow-up.

Conclusions

The Micrus ACT microcoil device is a feasible alternative platinum coil system, offering advantages over the currently available GDC device. The spherical coil offers excellent conformability and stability as an outer basket for aneurysm packing, and it seems particularly effective for small and wide-necked aneurysms. The rapid detachment mechanism is helpful, especially in large aneurysms. The Micrus coil device was safe and reliable in this series.

Acknowledgments

The authors would like to thank Julia Yarnold, FCR, MA, and Joan Rischmiller, RN, of the Neurovascular Research Unit, The Radcliffe Infirmary, Oxford, United Kingdom, for their help and support and Jan Newport for her assistance.

Dr A. J. Molyneux is a medical advisor to Micrus Corporation and has stock interest in the company. Dr D. B. Niemann has received educational grant support from Micrus Corporation and its distributors.

References

- Guglielmi G, Vinuela F, Sepetka I, et al. Electrothrombosis of saccular aneurysms via endovascular approach, part I: electrochemical basis technique and experimental results. J Neurosurg 1991;75:1–7
- Guglielmi G, Vinuela F, Dion J, et al. Electrothrombosis of saccular aneurysms via endovascular approach, II: preliminary clinical experience. J Neurosurg 1991;75:8–14
- Malek AM, Higashida RT, Phatouros CC, et al. Treatment of an intracranial aneurysm using a new three-dimensional-shape Guglielmi detachable coil, technical case report. *Neurosurgery* 1999; 44(45):1142–1144, discussion 1144–1145
- Drake CG. Report of World Federation of Neurological Surgeons Committee on a universal subarachnoid haemorrhage grading scale. J Neurosurg 1988;68:985–986
- Rankin J. Cerebrovascular accidents in patients over the age of 60: prognosis. Scott Med J 1957;2;2000–2015
- Lindley RI, Waddell F, Livingston M, et al. Can simple questions assess outcome after stroke? *Cerebrovasc Dis* 1994;4:314–324
- Molyneux AJ, Kerr RSCK, Stratton I., et al. International Subarachnoid Aneurysm Trial (ISAT) of neurosurgical clipping versus endovascular coiling in 2143 patients with ruptured intracranial aneurysms: a randomized trial. *Lancet* 2002;360:1267–1274
- Byrne JV, Molyneux AJ, Brennan RP, et al. Embolisation of recently ruptured intracranial aneurysms. J Neurol Neurosurg Psychiatry 1995;59:616–620
- 9. Cognard C, Weill A, Rey A, et al. Intracranial berry aneurysms: angiographic and clinical results after endovascular treatment. *Radiology* 1998;6:499–510
- Vinuela F, Duckwiler G, Mawaad M, Guglielmi G. Detachable coil embolisation of acute intracranial aneurysm, perioperative anatomical and clinical outcome in 403 patients. J Neurosurg 1997;86:475–482
- 11. The International Study of Unruptured Intracranial Aneurysms Investigators: unruptured intracranial aneurysms-risk of rupture and risks of surgical intervention. *N Engl J Med* 1998;339:1725–1733

818 NIEMANN

- 12. Johnston SC, Higashida RT, Barrow DL, et al. Recommendations for the endovascular treatment of intracranial aneurysms. A statement for healthcare professionals from the committee on the cerebrovascular imaging of the American Heart Association Council on cardiovascular radiology. *Stroke* 2002;33:2536–3544
- Fernandez Zubillagaq A, Guglielmi G, Vinuela F, Duckwiler GR. Endovascular occlusion of intracranial aneurysms with electrically detachable coils, correlation of aneurysm neck size and treatment

results. AJNR Am J Neuroradiol 1994;15:815-820

- Mawad ME, Cekirge S, Ciceri E, Saatci I. Endovascular treatment of giant and large intracranial aneurysms by using a combination of stent placement and liquid polymer injection. J Neurosurg 2002; 96(3):474-482
- Byrne JV, Sohn MJ, Molyneux AJ. Five-year experience in using coil embolisation for ruptured intracranial aneurysms, outcomes and incidence of late rebleeding. J Neurosurg 1999;90:656-663